

DECLARATION OF CONFORMANCE

It is hereby declared that the software:

- BKConnect Version 30.0.0.281 and FirmWare 3.5.0.2545
- PULSE Labshop Version 30.0.0.281 and FirmWare 3.5.0.2545

meets the requirements of the Standards:

- IEC 61672-1:2013, SLM Class 1
- IEC 61260-1:2014 Class 1 for Octave filters.

The performance is checked through our internal calibration procedures covering IEC61672-3 and IEC 61260-3 and is valid the following Microphone Combos and LAN-XI Data Acquisition Modules:

Microphones Combos (Microphone including preamplifiers):

- 4966-H-041 inclusive Preamplifier 1706
- 4189-A-021 inclusive Preamplifier 2671
- 4189-B/C/L-001 inclusive Preamplifier 2669 B/C/L
- Type 4189-H-041 inclusive Preamplifier 1706
- Type 4188-C/L-001 inclusive Preamplifier 2669 C/L
- 4188-A-021 inclusive Preamplifier 2671

LAN-XI Modules (all variants):

- 3160
- 3050
- 3053
- 3058
- 3676
- 3677

Date: August 10, 2026



Calibration Manager